

3697-3700

No.VMGMC/Store/Q/ /2018  
Office of the Dean,  
Dr. V. M. Govt. Medical College,  
Solapur Date: 09/03/2018

To

1. Health Diagnostic, Municipal Complex, 1<sup>st</sup> Floor, Main Road, Latur
2. Gunjana Surgical & Scientific Co., Mumbai
3. Rishi Surgical, Mumbai
4. N.P. Surgical Industries, Mumbai
5. Notice Board vmgmc, Solapur

**Sub:- Quotation of Instruments.**

Sealed quotations are invited for the following item from the stockiest for the use in various (Gynac.) department of Dr. V. M. Govt. Medical College, Solapur.

Name of Item	N.S.T. Machine Specifications
<b>N.S.T. Machine</b>	Size : 383mm(L) x 383 mm (W) x 124 mm (H)
	Net weight: <4 kg
	Display : 12.1IFT LCD
	Resolution: 800*600
	AC power: 100v – 200v 50hz/60hz
	DC power : 11.1v 4Ah rechargeable lithium battery
	Battery work time : >4 hours
	Environmental specification
	<b>Operation</b>
	Environmental temperature: +5c +40c
	Relative humidity : 25%- 93%
	Atmospheric pressure: 860 hPa – 1060 hPa
	Record
	Recording way: thermal printing system
	Recording paper: 152 mm (widel) x 13.5 (long)
	Paper type: Z-Food
	Paper speed : 1cm /min, 2 cm/ min, 3cm/min
	Resolution : 8dots/min
	<b>Alarm</b>
	Audible and Visible
	<b>Fetal Heart Rate (FHR) Measuring</b>
	Measuring Method : ultrasonic Doppler
	Measuring range: 50- 210 BPM
	Accuracy: +2BPM
	Ultrasonic frequency : 1Mhz
	Ultrasound output intensity: <5mW/cm2
	Display: wave/numerical
	<b>Environmental Specification</b>
Measuring method: External measurement	
Measuring range: 0-100 (%)	
Resolution : 1%	
Nonlinear error : <+10%	
<b>Fetal Movement recording</b>	
Manual button marking (MFM)	
Automatic FM identifying function (AFM)	
F80	

e:\2016-17 store\quotation (english)\dean level instrument purchase 17-18.docx

**Standard Configuration**

FHR TOCO FM Lithium battery printer

**Option**

TWIN FHR fetal stimulator CTG evolution module

Touch screen

**F 90 FETAL/MATERNAL MONITOR**

**Standard Configuration**

FHR, TOCO, FM Lithium battery printer

NIBP MSpoz MPR MECG MREST MTEMP

**Option**

TWIN FHR Fetal stimulator, CTG evaluation Module

Touch screen

The sealed quotations may be sent in the name of the Dean, Dr. V. M. Govt. Medical College, Solapur on or before **due date 16/3/2018**. Quotations are received after due date will not be considered.

Please subscribe the words on the top of envelop "Quotation for Instruments."

**Quotation Opening Date 19/3/2018**. Venue - College meeting hall.

  
**DEAN,**

DR. V. M. GOVT. MEDICAL COLLEGE,  
SOLAPUR